

F.A.Q
TENDER NO. 91180276

TENDER TITLE: Supply, Testing & Commissioning of Milk Analyzer Unit-100 Nos

DATE: April 13, 2024

SR. NO	QUERIES	RESPONSE BY GIZ
1	Any kind of Bid Security/ EMD and Tender Doc Fees Required by you for current tender?	NO
2	We are the manufacturer of EWS with brand name 'Akashganga' (with model approval) hence will supply our own make EWS with Similar Tech Specs as per your tender requirements instead of EWS-TI-10 Series. Pls Confirm?	Yes, you may quote for any brand name with the same specification.
3	Can we supply Milk Analyzer, Stirrer, EWS & Data Processor with a thermal printer as a separate component viz. interface together through suitable data cable/ wireless. Pls Confirm?	As long as all the appliances are compatible with each other, economical and easy to work with. Also, the maintenance contract should be the same for all of them.
4	What is your expected overall warranty period (in Yrs) decided for the current tender? And also advise AMC Process?	Overall warranty period – 1 year AMC – period of 2 years (at least)
5	What's the Total % of Perf Bank Guarantee & its period?	At the close of the work and before issue of final certificate of virtual completion, the contractor will furnish written performance guarantee against defective materials and workmanship for a period of one year from date of testing, commissioning, and satisfactory handing over. The Contractor will hold himself fully responsible for reinstallation or replacement, free of cost to Owner because of defective workmanship.
6	Delivery Period (In months)?	The equipment delivery must be coordinated with GIZ before the dispatch to the project site. The delivery period is 1.5 months (45 days from award of contract with complete Delivery, installation, trial runs of equipment and associated training at project site)
7	Pls specify Tech Specs of Solar Charge controller & Panel in detail!	Solar charge controller: A.C/D.C both Solar panel - Wattage: 100 Watt Brand: Exide make Controller: Imported Battery: Exide make, 80 Ah
8	Warranty of supply equipment?	Overall warranty period – 1 year
9	All required documents have to be sent through a given email only but if there limit of size of email to your email id then how can we send?	Please upload them on drive / SharePoint that is easily accessible to the email id mentioned in the Terms of Reference (TOR).
10	Please share the full specification of solar panel and battery for solar?	Solar charge controller: A.C/D.C both Solar panel - Wattage: 100 Watt Wattage: 100 Watt

F.A.Q
TENDER NO. 91180276

TENDER TITLE: Supply, Testing & Commissioning of Milk Analyzer Unit-100 Nos

DATE: April 13, 2024

		Brand: Exide make Controller: Imported Battery: Exide make, 80 Ah
11	Weighing Scale TI-10 Series mentioned in the inquiry which is some particular brand, we have our own brand of Prompt Make, can we quote our brand weighing scale? also confirm whether you want only TOP SS or under frame material as ss?	Weighing scale – Auto tare, Class III, configured with analyzer for online weighting data transfer analyzer. As long as the quality and specifications are the same, you may quote the other brand. Please make sure that it is compatible with other appliances in the setup. All steel construction with industrial coating.
